



PCR clean-up Gel extraction

User manual

NucleoSpin® Gel and PCR
Clean-up Midi

NucleoSpin® Gel and PCR
Clean-up Maxi

December 2019 / Rev. 06

Contact MN

Germany and international

MACHEREY-NAGEL GmbH & Co. KG

Neumann-Neander-Str. 6–8 · 52355 Düren · Germany

Tel.: +49 24 21 969-0

Toll-free: 0800 26 16 000 (Germany only)

Fax: +49 24 21 969-199

E-mail: info@mn-net.com

Technical Support Bioanalysis

Tel.: +49 24 21 969-270

E-mail: tech-bio@mn-net.com

USA

MACHEREY-NAGEL Inc.

2850 Emrick Blvd. · Bethlehem, PA 18020 · USA

Tel.: +1 484 821 0984

Toll-free: 888 321 6224 (MACH)

Fax: +1 484 821 1272

E-mail: sales-us@mn-net.com

France

MACHEREY-NAGEL SARL à associé unique

1, rue Gutenberg · 67722 Hoerdt · France

Tel.: +33 388 68 22 68

Fax: +33 388 51 76 88

E-mail: sales-fr@mn-net.com

Switzerland

MACHEREY-NAGEL AG

Hirsackerstr. 7 · 4702 Oensingen · Switzerland

Tel.: +41 62 388 55 00

Fax: +41 62 388 55 05

E-mail: sales-ch@mn-net.com

www.mn-net.com

Table of contents

1 Components	4
1.1 Kit contents	4
1.2 Reagents, consumables, and equipment to be supplied by user	5
2 Product description	6
2.1 The basic principle	6
2.2 Kit specifications	6
2.3 pH indicator	7
2.4 Tips and tricks for DNA extraction from agarose gels	8
2.5 DNA recovery depends on fragment size and elution volume	8
2.6 Salt carry-over and low A_{260} / A_{230}	10
3 Storage conditions and preparation of working solutions	12
4 Safety instructions	13
5 NucleoSpin® Gel and PCR Clean-up Midi protocols	14
5.1 PCR clean-up	14
5.2 DNA extraction from agarose gels	16
5.3 RNA extraction from agarose gels (Buffer NTC)	18
5.4 DNA clean-up of samples containing SDS (Buffer NTB)	19
5.5 Single stranded DNA clean-up (Buffer NTC)	20
6 NucleoSpin® Gel and PCR Clean-up Maxi protocols	21
6.1 PCR clean-up	21
6.2 DNA extraction from agarose gels	23
6.3 RNA extraction from agarose gels (Buffer NTC)	25
6.4 DNA clean-up of samples containing SDS (Buffer NTB)	26
6.5 Single stranded DNA clean-up (Buffer NTC)	27
7 Appendix	28
7.1 Troubleshooting	28
7.2 Ordering information	30
7.3 References	31
7.4 Product use restriction / warranty	32

1 Components

1.1 Kit contents

	NucleoSpin® Gel and PCR Clean-up Midi	NucleoSpin® Gel and PCR Clean-up Maxi
REF	20 preps 740986.20	20 preps 740610.20
Binding Buffer NT1	200 mL	450 mL
Wash Buffer NT3 (Concentrate)*	50 mL	100 mL
Elution Buffer NE**	30 mL	60 mL
NucleoSpin® Gel and PCR Clean-up Midi Columns	20	–
NucleoSpin® Gel and PCR Clean-up Maxi Columns	–	20
User manual	1	1

* For preparation of working solutions and storage conditions see section 3.

** Composition of Elution Buffer NE: 5 mM Tris/HCl, pH 8.5

1.2 Reagents, consumables, and equipment to be supplied by user

Reagents

- 96–100 % ethanol

Consumables

- 15 mL centrifuge tubes (NucleoSpin® Gel and PCR Clean-up Midi)
- 50 mL centrifuge tubes (NucleoSpin® Gel and PCR Clean-up Maxi)
- Disposable pipette tips

Equipment

- Manual pipettors
- Centrifuge for 15 mL tubes (NucleoSpin® Gel and PCR Clean-up Midi)
- Centrifuge for 50 mL tubes (NucleoSpin® Gel and PCR Clean-up Maxi)
- Heating block, water bath, or thermomixer for gel extraction
- Scalpel to cut agarose gels
- Vortex mixer
- Personal protection equipment (lab coat, gloves, goggles)

2 Product description

2.1 The basic principle

NucleoSpin® Gel and PCR Clean-up Midi and **NucleoSpin® Gel and PCR Clean-up Maxi** are developed as 2 in 1 kits. DNA fragments can be purified from enzymatic reactions such as PCR as well as from agarose gels.

The sample is mixed with Binding Buffer NT1 and, in case of a cut out gel band, heated to dissolve the agarose. In the presence of chaotropic salt the DNA is bound to the silica membrane of a NucleoSpin® Gel and PCR Clean-up Column. Contaminations are removed by simple washing steps with ethanolic Wash Buffer NT3. Pure DNA is finally eluted under low salt conditions with slightly alkaline Elution Buffer NE (5 mM Tris/HCl, pH 8.5).

2.2 Kit specifications

- **NucleoSpin® Gel and PCR Clean-up kits** are designed for fast purification of PCR products or DNA from enzymatic reactions, and the extraction of DNA fragments from TAE or TBE agarose gels.
- The **NucleoSpin® Gel and PCR Clean-up Midi** allows the processing of up to 4 mL of PCR/enzymatic reaction or 4 g of agarose gel.
- The **NucleoSpin® Gel and PCR Clean-up Maxi** allows the processing of up to 10 mL of PCR/enzymatic reaction or 10 g of agarose gel.
- The **NucleoSpin® Gel and PCR Clean-up** buffer formulations ensures complete removal of all kinds of contaminations like
 - nucleotides, primers
 - enzymes
 - mineral oil
 - PCR additives (e.g., salts, betaine, DMSO)
 - detergents (e.g., Tween 20, Triton X-100)
 - dyes (e.g., ethidiumbromide)
 - unbound labels and tags
- Primers from PCR reactions are quantitatively eliminated while small DNA fragments are still bound and purified with high recovery.
- **NucleoSpin® PCR and Gel Clean-up kits** can be used with all kinds of agarose gels (high or low melting) with 1 % to 5 % agarose and a variety of buffer systems like TAE or TBE (tips and tricks in section 2.4).
- Several support protocols extend the application range of **NucleoSpin® Gel and PCR Clean-up kits** to
 - Clean-up of DNA from reaction mixtures containing SDS (section 5.4, 6.4)
 - Clean-up of single stranded DNA (section 5.5, 6.5)
 - Extraction of RNA from agarose gels (section 5.3, 6.3)

- The purified and concentrated DNA can directly be used for hybridization, sequencing, PCR, restriction, ligation, in vitro transcription, labeling or any other kind of enzymatic reaction.


Table 1: Kit specifications at a glance

Parameter	NucleoSpin® Gel and PCR Clean-up Midi	NucleoSpin® Gel and PCR Clean-up Maxi
Sample material	Up to 4 mL of PCR reaction or 4 g of gel	Up to 10 mL of PCR reaction or 10 g of gel
Binding capacity	75 µg	250 µg
Fragment length	50 bp– ~ 20 kbp	50 bp– ~ 20 kbp
Elution volume	200–400 µL	1000 µL

2.3 pH indicator

The optimal pH to bind even small DNA fragments to the silica membrane of the NucleoSpin® Gel and PCR Clean-up Column is approximately 5.0–6.0. The Binding Buffer NTI is sufficiently buffered to maintain this pH for all standard PCR reaction buffers or agarose gel buffer systems.

In addition, the colored binding buffer helps identify undissolved pieces of agarose during DNA gel extraction.


Figure 1 Titration curve of Binding Buffer NTI with pH indicator

A yellow color indicates the optimal pH < 6.0 (Figure 1). If the pH increases to around 7 after adding the sample, the solution will turn green. In case the pH is higher than 7 the solution turns blue. If a change in color is observed, the pH should be adjusted by adding more Buffer NTI or by titrating the pH to < 6.0 with 4 M sodium acetate pH 5.0 or small amounts of hydrochloric acid (HCl).

2.4 Tips and tricks for DNA extraction from agarose gels

Subject	Recommendation
Buffer system	TBE (Tris-Borate-EDTA) buffer has a higher buffering capacity than TAE (Tris-Acetate-EDTA) which is needed for runs overnight and offers a better resolution for small DNA fragments. TBE buffer can be used in combination with NucleoSpin® Gel and PCR Clean-up kits. However, fresh TAE buffer should be preferred over TBE for preparative agarose gels. TAE does not interact with agarose resulting in higher DNA yields. Additionally, linear DNA runs faster and the resolution of large DNA fragments is higher. Furthermore, supercoiled plasmid is separated better from linear and open circle DNA.
Running conditions	The temperature during electrophoresis should be low to increase the resolution of the DNA separation and to avoid melting of the gel and denaturation of DNA. Use fresh buffer, run the gel at low voltage (< 60 V), and as short as possible. As soon as the DNA band of interest is sufficiently separated from the rest, stop the gel and cut out the band.
Cutting out the band	Expose the gel to UV light as short as possible. Use the longest UV wave length that is allowed by your gel documentation system. Prolonged exposure and short wave lengths can damage the DNA. Wear gloves and a face mask to protect your skin and eyes from UV light. Make sure to cut through the gel vertically and remove all excess agarose. Use 0.7–1.0 % agarose gels rather than higher percentages.
Size of gel piece	Make sure to actually weigh the gel since its weight is easily underestimated. Up to 10 g of agarose gel can be dissolved with 10 mL of Buffer NT1 and loaded onto the column in two steps. Please note that NucleoSpin® Gel and PCR Clean-up Midi is limited to 4g of agarose gel.

2.5 DNA recovery depends on fragment size and elution volume

After washing with Buffer NT3 the DNA is sticking to the silica membrane. The number of interactions with Si-OH groups of the silica increase with the size of the DNA fragment. Therefore, large DNA with several kilo base pairs binds much stronger and is much more difficult to elute than small DNA with just several hundred base pairs. **NucleoSpin® Gel and PCR Clean-up kits** are recommended for DNA up to 10–15 kbp. Longer fragments can be purified, but recovery might be low. Furthermore, fragments larger than 20 kbp might be mechanically damaged by the fast centrifugation through the membrane. For very large fragments using NucleoTrap® or NucleoTrap®CR should be considered (see ordering information, 7.2).

To elute the DNA, water and a pH > 7 is needed to reestablish the hydrate shell. It is highly recommended to **elute DNA with Elution Buffer NE** (5 mM Tris/HCl, pH 8.5) provided with the kit or standard TE buffer to ensure best elution efficiency. Note that EDTA in TE buffer may cause problems in subsequent enzymatic reactions. Do not use deionized water since its pH is usually too acidic. If even less salt than 5 mM Tris has to be used, dilute Elution Buffer NE with distilled water and make sure, the pH is still > 7. Unbuffered elution buffer should not be used.

The standard elution buffer volume is 2 x 200 µL (NucleoSpin® Gel and PCR Clean-up Midi) or **1000 µL** (NucleoSpin® Gel and PCR Clean-up Maxi) which is the best compromise for high DNA recovery and high DNA concentration for fragments < 1000 bp.

Elution after gel extraction is 10–20 % less efficient than elution of purified PCR products. Elution of several kbp long DNA fragments is 10–30 % less efficient than elution of 500 bp fragments. To improve the DNA recovery after gel extraction and/or for large DNA fragments, the following modifications can be applied to the standard elution procedure:

- Heat elution buffer to 70 °C and incubate elution buffer on the column at 70 °C for 5 minutes.
- Apply elution buffer to the column and centrifuge first at 30–50 x g for 1 min and then at 11.000 x g for 1 min.
- Do **2 or better 3 elution steps with 200 µL** (NucleoSpin® Gel and PCR Clean-up Midi) or **1000 µL** (NucleoSpin® Gel and PCR Clean-up Maxi) **with the same elution buffer.**

2.6 Salt carry-over and low A_{260}/A_{230}

The silica membrane technology to purify RNA or DNA is based on the ability of chaotropic salts to destroy the water shell around nucleic acids. Two commonly used chaotropic salts are guanidine hydrochloride (GuHCl) and guanidinium thiocyanate (GuSCN). In solution they both have the same guanidinium cation but different anions. These anions are not only responsible for their different behavior towards nucleic acids but also for their different UV absorption spectra. GuHCl exhibits only minimal absorption < 220 nm even at a concentration of 1 M, whereas GuSCN already shows significant absorption < 240 nm (1 mM, Figure 1) and even < 260 nm (1 M).




Figure 2 UV absorption spectra of 1 mM GuHCl (solid line) and 1 mM GuSCN (dotted line)

Especially the difference in absorption at 230 nm can have a huge impact on the purity ratio A_{260}/A_{230} if DNA is contaminated with chaotropic salts. Carry over of GuSCN can lower the ratio from its ideal value of > 2.0 to below 1.5 or even 1.0. GuHCl on the other hand is invisible at this wave length and does not alter the ratio at all. This effect, however, is only detectable with very small amounts of DNA like typical yields of PCR reactions or gel extractions. Technical advances in UV-VIS spectrometry now allow measuring these small amounts of DNA in small volumes thereby raising concerns that the DNA might be too “dirty”. This problem does usually not occur with larger amounts of DNA since its own absorption at 230 nm masks the small contribution of any contamination.

The concentration of contaminating chaotropic salt is usually in the range of 100 μ M to 1 mM and does not have any negative influence on enzymatic downstream applications like, e.g., PCR, restriction, or ligation. Figure 2 shows qPCR inhibition by GuSCN and GuHCl. It demonstrates that PCR only starts to be inhibited by chaotropic salts with about 100-fold higher concentration (40 mM). Furthermore it shows clearly, that not only GuSCN can cause inhibition but also the photometrically invisible GuHCl.




Figure 3 qPCR inhibition by GuHCl (light gray) and GuSCN (dark gray)

A 164 bp DNA fragment was amplified from 5 ng pBS template with DyNAmo Capillary Master Mix (NEB) in a Lighcycler real-time PCR machine (Roche) in the presence of 0–80 mM GuHCl or GuSCN.

Salt carry over always happens – with both GuSCN and GuHCl – and could only be minimized by most extensive washing. This, however, is unnecessary, since the final concentration of chaotropic salt in eluates is much too small to have any negative effect. Thus, non-ideal A_{260}/A_{230} can simply be ignored.

3 Storage conditions and preparation of working solutions

Attention: Buffer NT1 contains chaotropic salt. Wear gloves and goggles!

Storage conditions:

- All kit components can be stored at room temperature and are stable for at least one year.

Before starting any **NucleoSpin® Gel and PCR Clean-up** protocol prepare the following:

- Wash Buffer NT3:** Add the indicated volume of ethanol (96–100 %) to **Buffer NT3 Concentrate**. Mark the label of the bottle to indicate that ethanol was added. Wash Buffer NT3 is stable at room temperature (18–25 °C) for at least one year.

	NucleoSpin® Gel and PCR Clean-up Midi	NucleoSpin® Gel and PCR Clean-up Maxi
REF	20 preps 740986.20	20 preps 740610.20
Wash Buffer NT3 (Concentrate)	50 mL Add 200 mL ethanol	100 mL Add 400 mL ethanol

4 Safety instructions

The following components of the **NucleoSpin® Gel and PCR Clean-up Midi** and the **NucleoSpin® Gel and PCR Clean-up Maxi** kit contain hazardous contents. *Wear gloves and goggles and follow the safety instructions given in this section.*

Only harmful features do not need to be labeled with H and P phrases up to 125 mL or 125 g. *Mindergefährliche Eigenschaften müssen bis 125 mL oder 125 g nicht mit H- und P-Sätzen gekennzeichnet werden.*

Component	Hazard contents	GHS symbol	Hazard phrases	Precaution phrases
Inhalt	Gefahrstoff	GHS-Symbol	H-Sätze	P-Sätze
NTI	Guanidinium thiocyanate 30–60 % <i>Guanidinthiocyanat 30–45 %</i> CAS 593-84-0	 WARNING <i>ACHTUNG</i>	302, 412	264W, 273, 301+312, 330

Hazard phrases

- H 302 Harmful if swallowed.
Gesundheitsschädlich bei Verschlucken.
- H 412 Harmful to aquatic life with long lasting effects.
Schädlich für Wasserorganismen, mit langfristiger Wirkung.

Precaution phrases

- P 264W Wash with water thoroughly after handling.
Nach Gebrauch mit Wasser gründlich waschen.
- P 273 Avoid release to the environment.
Freisetzung in die Umwelt vermeiden.
- P 301+312 IF SWALLOWED: Call a POISON CENTER/ doctor/ ... / if you feel unwell.
BEI VERSCHLUCKEN: Bei Unwohlsein GIFTINFORMATIONSZENTRUM/Arzt/... anrufen.
- P 330 Rinse mouth.
Mund ausspülen.

For further information please see Material Safety Data Sheets (www.mn-net.com).
Weiterführende Informationen finden Sie in den Sicherheitsdatenblättern (www.mn-net.com).



The symbol shown on labels refers to further safety information in this section.

Das auf Etiketten dargestellte Symbol weist auf weitere Sicherheitsinformationen dieses Kapitels hin.

5 NucleoSpin® Gel and PCR Clean-up Midi protocols

5.1 PCR clean-up

The following protocol is suitable for PCR clean-up as well as DNA concentration and removal of salts, enzymes, etc. from enzymatic reactions (SDS < 0.1 %).

Before starting the preparation:

- Check if Wash Buffer NT3 was prepared according to section 3.

1 Adjust DNA binding condition

For sample volumes < 1 mL adjust the volume of the reaction mixture to 1 mL with water. It is not necessary to remove mineral oil.



**+ 2 vol NTI per
1 vol sample**

Mix **1 volume of sample** with **2 volumes of Buffer NTI** (e.g., mix 1 mL PCR reaction and 2 mL Buffer NTI).

2 Bind DNA

Place a **NucleoSpin® Gel and PCR Clean-up Midi Column** into a Collection Tube (15 mL) and load the sample.



Load sample

Centrifuge for **1 min at $\geq 3,000 \times g$** . Discard flow-through and place the column back into the collection tube.



**$\geq 3,000 \times g$
1 min**

3 Wash silica membrane

Add **4 mL Buffer NT3** to the NucleoSpin® Gel and PCR Clean-up Midi Column. Centrifuge for **1 min at $\geq 3,000 \times g$** . Discard flow-through and place the column back into the collection tube.



+ 4 mL NT3



**$\geq 3,000 \times g$
1 min**

Repeat the washing step.



+ 4 mL NT3



**$\geq 3,000 \times g$
1 min**

4 Dry silica membrane

Centrifuge for **10 min** at **$\geq 3,000 \times g$** to remove **Buffer NT3** completely. Make sure the spin column does not come in contact with the flow through while removing it from the centrifuge and the collection tube.



Note: Residual ethanol from Buffer NT3 might inhibit enzymatic reactions. Total removal of ethanol can be achieved by incubating the columns for 2–5 min at 70 °C prior to elution.



**$\geq 3,000 \times g$
10 min**

5 Elute DNA

Place the NucleoSpin® Gel and PCR Clean-up Midi Column into a **new** 15 mL centrifuge tube (not provided). Add **200 μ L Buffer NE** and incubate at **70 °C** for **5 min**. Centrifuge for **2 min** at **$\geq 3,000 \times g$** .



+ 200 μ L NE
70 °C
5 min



**$\geq 3,000 \times g$
2 min**

Repeat elution with fresh elution buffer.



+ 200 μ L NE
70 °C
5 min



**$\geq 3,000 \times g$
2 min**

5.2 DNA extraction from agarose gels


Before starting the preparation:

- Check if Wash Buffer NT3 was prepared according to section 3.

1 Excise DNA fragment/solubilize gel slice

Note: Minimize UV exposure time to avoid damaging the DNA. Refer to section 2.4 for more tips on agarose gel extraction.

Take a clean scalpel to excise the DNA fragment from an agarose gel. Remove all excess agarose.



Determine the weight of the gel slice and transfer it to a clean tube.

For each **1 g of agarose gel < 2 %** add **1 mL Buffer NTI**.

For gels containing **> 2 %** agarose, double the volume of Buffer NTI.

Incubate sample for **10–20 min** at **50 °C**. Vortex the sample briefly every 2–3 min until the gel slice is **completely** dissolved!



**+ 1 mL NTI per
1 g gel**

**50 °C
10–20 min**

2 Bind DNA

Place a **NucleoSpin® Gel and PCR Clean-up Midi Column** into a Collection Tube (15 mL) and load the sample.

Centrifuge for **1 min** at **≥ 3,000 x g**. Discard flow through and place the column back into the collection tube.

Repeat this step to load more sample material.



Load sample



**≥ 3,000 x g
1 min**

3 Wash silica membrane

Add **4 mL Buffer NT3** to the NucleoSpin® Gel and PCR Clean-up Midi Column. Centrifuge for **1 min** at **≥ 3,000 x g**. Discard flow through and place the column back into the collection tube.



+ 4 mL NT3



≥ 3,000 x g
1 min

Repeat this washing step.



+ 4 mL NT3



≥ 3,000 x g
1 min

4 Dry silica membrane

Centrifuge for **10 min** at **≥ 3,000 x g** to remove **Buffer NT3** completely. Make sure the spin column does not come in contact with the flow through while removing it from the centrifuge and the collection tube.



***Note:** Residual ethanol from Buffer NT3 might inhibit enzymatic reactions. Total removal of ethanol can be achieved by incubating the columns for 2–5 min at 70 °C prior to elution.*



≥ 3,000 x g
10 min

5 Elute DNA

Place the NucleoSpin® Gel and PCR Clean-up Midi Column into a **new** 15 mL centrifuge tube (not provided). Add **200 µL Buffer NE** and incubate at **70 °C** for **5 min**. Centrifuge for **2 min** at **≥ 3,000 x g**.



+ 200 µL NE
70 °C
5 min



≥ 3,000 x g
2 min

Repeat elution with fresh elution buffer.



+ 200 µL NE
70 °C
5 min



≥ 3,000 x g
2 min

5.3 RNA extraction from agarose gels (Buffer NTC)

Not only DNA but also RNA can be extracted from agarose gels. To efficiently bind especially the small, single stranded RNA **Binding Buffer NTC** has to be used instead of standard Binding Buffer NTI.

To fractionate RNA run a standard RNA gel with denaturing RNA loading buffer but **do not use formaldehyde or glyoxal**. These compounds not only inactivate RNases and denature RNA but also modify RNA. As a result the RNA yield is significantly reduced and more important the RNA might not work properly in enzymatic downstream applications like RT-PCR or *in vitro* transcriptions.

Without formaldehyde the RNA is very sensitive to contaminating RNases. Use gloves and make sure all equipment is RNase-free, especially the agarose, and the running buffers. Run the gel as short and as cold (low voltage, cold room) as possible. Note that the RNA might form secondary structures and might run differently from denaturing agarose gels.


Note: Buffer NTC has to be ordered separately (125 mL Buffer NTC, REF 740654.100, see ordering information, 7.2)

Before starting the preparation:

- Check if Wash Buffer NT3 was prepared according to section 3.

1 Excise DNA fragment / solubilize gel slice

Note: Minimize UV exposure time to avoid damaging the DNA. Refer to section 2.4 for more tips on agarose gel extraction.



Take a clean scalpel to excise the DNA fragment from an agarose gel. Remove all excess agarose.



Determine the weight of the gel slice and transfer it to a clean tube.

For each **1 g of agarose gel < 2 %** add **1 mL Buffer NTC**.

For gels containing **> 2 %** agarose, double the volume of Buffer NTC.

Incubate sample for **10–20 min** at **50 °C**. Vortex the sample briefly every 2–3 min until the gel slice is **completely** dissolved!



**+ 1 vol NTC
per 1 vol
sample**

**50 °C
10–20 min**

2 Bind RNA

Continue with **step 2** of the protocol for DNA extraction from agarose gels (section 5.2).

5.4 DNA clean-up of samples containing SDS (Buffer NTB)

Buffer NTI from the NucleoSpin® Gel and PCR Clean-up kit is compatible with most commonly used detergents except sodium dodecyl sulfate (SDS). For purification of DNA from samples without SDS the standard protocol for PCR clean-up can be used (see section 5.1). For purification of DNA from SDS containing buffers, for example in applications like “Chromatin Immunoprecipitation” (ChIP), the SDS compatible Binding Buffer NTB can be used.

Note: Buffer NTB has to be ordered separately (150 mL Buffer NTB, REF 740595.150, see ordering information, section 7.2).

Before starting the preparation:

- Check if Wash Buffer NT3 was prepared according to section 3.

1 Adjust DNA binding condition

Mix **1 volume of sample** with **5 volumes of Buffer NTB**
(e.g., 1 mL reaction mix with 5 mL Buffer NTB).

**+ 5 vol NTB
per
1 vol sample**

Note: If SDS starts to precipitate add 1 volume of isopropanol or warm sample to 20–30 °C.

2 Bind DNA

Continue with **step 2** of the protocol for PCR clean up
(section 5.1).

5.5 Single stranded DNA clean-up (Buffer NTC)

Buffer NTI from the NucleoSpin® Gel and PCR Clean-up kit is able to bind single stranded DNA (ssDNA) > 150 bases. Shorter oligonucleotides, especially primers, are completely removed. If you need to purify short ssDNA the additional Binding Buffer NTC can be used (see Figure 4).

Note: Buffer NTC has to be ordered separately (125 mL Buffer NTC, REF 740654.100, see ordering information, section 7.2).

Before starting the preparation:

- Check if Wash Buffer NT3 was prepared according to section 3.

1 Adjust DNA binding condition

Mix **1 volume of sample** with **2 volumes** of **Buffer NTC**
(e.g., 1 mL PCR reaction mix and 2 mL Buffer NTC).

**+ 2 vol NTC
per
1 vol sample**

If your sample contains large amounts of detergents or other critical substances, double the volume of Buffer NTC.

2 Bind DNA

Continue with **step 2** of the protocol for PCR clean up (section 5.1).




Figure 4 Purification of dsDNA and ssDNA using buffers NTI and NTC

PCR fragments, amplified using one phosphorylated and one dephosphorylated primer, were partially digested with λ -Exonuclease to yield single stranded DNA. Samples were purified using Binding Buffer NTI and NTC and run on a 1 % TAE agarose gel. Remaining double stranded DNA can be seen as faint bands. The corresponding single stranded DNA is running slightly faster due to secondary structure formation. Compared to the input DNA (u, lane 1), Buffer NTI removes ssDNA < 150 bases (NTI, lane 2), whereas Buffer NTC leads to full recovery of even primer oligonucleotides (NTC, lane 3).

6 NucleoSpin® Gel and PCR Clean-up Maxi protocols

6.1 PCR clean-up

The following protocol is suitable for PCR clean-up as well as DNA concentration and removal of salts, enzymes, etc. from enzymatic reactions (SDS < 0.1 %).

Before starting the preparation:

- Check if Wash Buffer NT3 was prepared according to section 3.

1 Adjust DNA binding condition

For sample volumes < 1 mL adjust the volume of the reaction mixture to 1 mL with water. It is not necessary to remove mineral oil.

Mix **1 volume of sample** with **2 volumes of Buffer NT1** (e.g., mix 1 mL PCR reaction and 2 mL Buffer NT1).



**+ 2 vol NT1 per
1 vol sample**

2 Bind DNA

Place a **NucleoSpin® Gel and PCR Clean-up Maxi Column** into a Collection Tube (50 mL) and load the sample.

Centrifuge for **1 min at $\geq 3,000 \times g$** . Discard flow-through and place the column back into the collection tube.



Load sample



**$\geq 3,000 \times g$
1 min**

3 Wash silica membrane

Add **10 mL Buffer NT3** to the NucleoSpin® Gel and PCR Clean-up Maxi Column. Centrifuge for **1 min at $\geq 3,000 \times g$** . Discard flow through and place the column back into the collection tube.



+ 10 mL NT3



**$\geq 3,000 \times g$
1 min**



+ 10 mL NT3



**$\geq 3,000 \times g$
1 min**

Repeat the washing step.

4 Dry silica membrane

Centrifuge for **10 min** at $\geq 3,000 \times g$ to remove **Buffer NT3** completely. Make sure the spin column does not come in contact with the flow-through while removing it from the centrifuge and the collection tube.



Note: Residual ethanol from Buffer NT3 might inhibit enzymatic reactions. Total removal of ethanol can be achieved by incubating the columns for 2–5 min at 70 °C prior to elution.



$\geq 3,000 \times g$
10 min

5 Elute DNA

Place the NucleoSpin® Gel and PCR Clean-up Midi Column into a **new** 50 mL centrifuge tube (not provided). Add **1000 µL Buffer NE** and incubate at **70 °C** for **5 min**. Centrifuge for **2 min** at $\geq 3,000 \times g$.



+ 1000 µL NE
70 °C
5 min



$\geq 3,000 \times g$
2 min

Reload eluate into the column and repeat elution at least once.

Note: Repeated elution steps with the same volume of Buffer NE will increase the DNA recovery without reducing the DNA concentration.



**Reload eluate
into column**
70 °C
5 min

$\geq 3,000 \times g$
2 min

6.2 DNA extraction from agarose gels


Before starting the preparation:

- Check if Wash Buffer NT3 was prepared according to section 3.

1 Excise DNA fragment/solubilize gel slice

Note: Minimize UV exposure time to avoid damaging the DNA. Refer to section 2.4 for more tips on agarose gel extraction.

Take a clean scalpel to excise the DNA fragment from an agarose gel. Remove all excess agarose.



Determine the weight of the gel slice and transfer it to a clean tube.

For each **1 g of agarose gel < 2 %** add **1 mL Buffer NTI**.

For gels containing **> 2 %** agarose, double the volume of Buffer NTI.

Incubate sample for **10–20 min** at **50 °C**. Vortex the sample briefly every 2–3 min until the gel slice is **completely** dissolved!



**+ 1 mL NTI per
1 g gel**

**50 °C
10–20 min**

2 Bind DNA

Place a **NucleoSpin® Gel and PCR Clean-up Maxi Column** into a Collection Tube (50 mL) and load the sample.

Centrifuge for **1 min** at **≥ 3,000 x g**. Discard flow through and place the column back into the collection tube.

Repeat this step to load more sample material.



Load sample



**≥ 3,000 x g
1 min**

3 Wash silica membrane

Add **10 mL Buffer NT3** to the NucleoSpin® Gel and PCR Clean-up Maxi Column. Centrifuge for **1 min** at **≥ 3,000 x g**. Discard flow through and place the column back into the collection tube.

Repeat this washing step.



+ 10 mL NT3



**≥ 3,000 x g
1 min**



+ 10 mL NT3



**≥ 3,000 x g
1 min**

4 Dry silica membrane

Centrifuge for **10 min** at **$\geq 3,000 \times g$** to remove **Buffer NT3** completely. Make sure the spin column does not come in contact with the flow through while removing it from the centrifuge and the collection tube.



Note: Residual ethanol from Buffer NT3 might inhibit enzymatic reactions. Total removal of ethanol can be achieved by incubating the columns for 2–5 min at 70 °C prior to elution.



**$\geq 3,000 \times g$
10 min**

5 Elute DNA

Place the NucleoSpin® Gel and PCR Clean-up Maxi Column into a **new** 50 mL centrifuge tube (not provided). Add **1000 μ L Buffer NE** and incubate at **70 °C** for **5 min**. Centrifuge for **2 min** at **$\geq 3,000 \times g$** .



+ 1000 μ L NE
70 °C
5 min



**$\geq 3,000 \times g$
2 min**

Reload eluate into the column and repeat elution at least once.

Note: Repeated elution steps with the same volume of Buffer NE will increase the DNA recovery without reducing the DNA concentration.



**Reload eluate
into column**
70 °C
5 min

**$\geq 3,000 \times g$
2 min**

6.3 RNA extraction from agarose gels (Buffer NTC)

Not only DNA but also RNA can be extracted from agarose gels. To efficiently bind especially the small, single stranded RNA **Binding Buffer NTC** has to be used instead of standard Binding Buffer NTI.

To fractionate RNA run a standard RNA gel with denaturing RNA loading buffer but **do not use formaldehyde or glyoxal**. These compounds not only inactivate RNases and denature RNA but also modify RNA. As a result the RNA yield is significantly reduced and more important the RNA might not work properly in enzymatic downstream applications like RT-PCR or *in-vitro* transcriptions.

Without formaldehyde the RNA is very sensitive to contaminating RNases. Use gloves and make sure all equipment is RNase-free, especially the agarose, and the running buffers. Run the gel as short and as cold (low voltage, cold room) as possible. Note that the RNA might form secondary structures and might run differently from denaturing agarose gels.

Note: Buffer NTC has to be ordered separately (125 mL Buffer NTC, REF 740654.100, see ordering information, section 7.3)

Before starting the preparation:

- Check if Wash Buffer NT3 was prepared according to section 3.

1 Adjust RNA binding condition

Mix **1 volume** of sample with **1 volume** of **Buffer NTC**
(e.g., 2 g gel and 2 mL Buffer NTC).

**+ 1 vol NTC
per 1 vol
sample**

2 Bind RNA

Continue with **step 2** of the protocol for DNA extraction
from agarose gels (section 6.2).

6.4 DNA clean-up of samples containing SDS (Buffer NTB)

Buffer NTI from the NucleoSpin® Gel and PCR Clean-up kit is compatible with most commonly used detergents except sodium dodecyl sulfate (SDS). For purification of DNA from samples without SDS the standard protocol for PCR clean-up can be used (see section 6.1). For purification of DNA from SDS containing buffers, for example in applications like “Chromatin Immunoprecipitation” (ChIP), the SDS compatible Binding Buffer NTB can be used.

Note: Buffer NTB has to be ordered separately (150 mL Buffer NTB, REF 740595.150, see ordering information, section 7.3).

Before starting the preparation:

- Check if Wash Buffer NT3 was prepared according to section 3.

1 Adjust DNA binding condition

Mix **1 volume of sample** with **5 volumes of Buffer NTB**
(e.g., 1 mL reaction mix with 5 mL Buffer NTB).

**+ 5 vol NTB
per
1 vol sample**

Note: If SDS starts to precipitate add 1 volume of isopropanol or warm sample to 20–30 °C.

2 Bind DNA

Continue with **step 2** of the protocol for PCR clean up
(section 6.1).

6.5 Single stranded DNA clean-up (Buffer NTC)

Buffer NTI from the NucleoSpin® Gel and PCR Clean-up Maxi kit is able to bind single stranded DNA (ssDNA) > 150 bases. Shorter oligonucleotides, especially primers, are completely removed. If you need to purify short ssDNA the additional Binding Buffer NTC can be used (see Figure 5).

Note: Buffer NTC has to be ordered separately (125 mL Buffer NTC, REF 740654.100, see ordering information, section 7.3).

Before starting the preparation:

- Check if Wash Buffer NT3 was prepared according to section 3.

1 Adjust DNA binding condition

Mix **1 volume of sample** with **2 volumes** of **Buffer NTC**
(e.g., 1 mL PCR reaction mix and 2 mL Buffer NTC).

**+ 2 vol NTC
per
1 vol sample**

If your sample contains large amounts of detergents or other critical substances, double the volume of Buffer NTC.

2 Bind DNA

Continue with **step 2** of the protocol for PCR clean up (section 6.1).




Figure 5 Purification of dsDNA and ssDNA using buffers NTI and NTC

PCR fragments, amplified using one phosphorylated and one dephosphorylated primer, were partially digested with λ -Exonuclease to yield single stranded DNA. Samples were purified using Binding Buffer NTI and NTC and run on a 1 % TAE agarose gel. Remaining double stranded DNA can be seen as faint bands. The corresponding single stranded DNA is running slightly faster due to secondary structure formation. Compared to the input DNA (u, lane 1), Buffer NTI removes ssDNA < 150 bases (NTI, lane 2), whereas Buffer NTC leads to full recovery of even primer oligonucleotides (NTC, lane 3).

7 Appendix

7.1 Troubleshooting

Problem	Possible cause and suggestions
Incomplete melting of agarose slices	<i>Time and temperature</i>
	Check incubation temperature and volume of Buffer NT1. Increase incubation time. Vortex every 2 min and check integrity of the gel slice. Very large gel slices can be crushed before addition of Buffer NT1 to shorten the melting time.
	<i>Reagents not prepared properly</i>
Low DNA yield	Add indicated volume of 96–100 % ethanol to Buffer NT3 Concentrate and mix well before use.
	<i>Incompletely dissolved gel slice</i>
	Increase time or add another two volumes of Buffer NT1 and vortex the tube every 2 minutes during incubation at 50 °C. Small pieces of gel are hardly visible and contain DNA that will be lost for purification.
	<i>Insufficient drying of the NucleoSpin® Gel and PCR Clean-up kit silica membrane</i>
	Centrifuge 10 min at $\geq 3,000 \times g$ or incubate column for 2–5 min at 70 °C before elution to remove ethanolic Buffer NT3 completely. Ethanolic contaminations are also indicated by gel-loading problems (samples float out of gel slots). Remove the spin cup carefully from the centrifuge and collection tube and avoid contact of spin cup with flow-through.
	<i>Incomplete elution</i>
	Especially for larger amounts of DNA, long DNA fragments (> 1000 bp), or after gel extraction, do multiple elution steps with fresh buffer, heat to 70 °C, and incubate for 5 min. See section 2.4 for detailed information.

Problem	Possible cause and suggestions
Appearance of additional bands on agarose gel after gel extraction	<i>DNA was denatured during purification</i> - In case water is used for elution and agarose with a low ion content is used for agarose gel electrophoresis, the formation of denatured (single-stranded) DNA might be promoted. To re-anneal the DNA, add all components of the subsequent enzymatic reaction omitting the enzyme. Incubate at 95 °C for 2 min and let the mixture cool slowly to room temperature (at this step the DNA re-anneals). Add the enzyme and continue with your downstream application. - Use fresh running buffer and run at low voltage to lower the temperature. High temperature might promote DNA denaturation during electrophoresis.
	<i>Carry over of ethanol/ethanolic Buffer NT3</i> - Centrifuge 10 min at $\geq 3,000 \times g$ or better incubate column for 5–10 min at 70 °C before elution to remove ethanolic Buffer NT3 completely. Ethanolic contaminations are also indicated by gel-loading problems (samples float out of gel slots). Remove the spin cup carefully from the centrifuge and collection tube, and avoid contacting the spin cup with the flow-through. - Use a different brand of ethanol to reconstitute Buffer NT3 or ethanol that is not denatured. The denaturing components might not evaporate as fast as ethanol, end up concentrated in the eluate, and inhibit enzymes like ligase.
Suboptimal performance of DNA in sequencing, restriction, or ligation reactions	<i>Carry over of chaotropic salts</i> - Perform the optional washing step. - Additionally, 1 mL NT3 can be loaded before the drying step. (Note: The volume of Buffer NT3 included in the kit is not sufficient for this modification for all preparations, but can be ordered separately, see ordering information, page 30.)
	<i>Elution of DNA with buffers other than Buffer NE, for example TE buffer (Tris/EDTA)</i> - EDTA might inhibit sequencing reactions. In this case it is recommended to re-purify DNA and elute in Buffer NE or water. <i>Not enough DNA used for sequencing reaction</i> - Quantify DNA by agarose gel electrophoresis before setting up sequencing reactions.

Problem	Possible cause and suggestions
Suboptimal performance of DNA in sequencing, restriction, or ligation reactions (continued)	<i>DNA was damaged by UV light</i> Reduce UV exposure time to a minimum when excising a DNA fragment from an agarose gel.
Suboptimal performance of DNA in NanoDrop® Spectrophotometer Analysis or Agilent's Bioanalyzer	<i>Carry-over of traces of silica particles</i> NanoDrop® Spectrophotometer technology is very sensitive to any particles included in the sample material. To pellet the silica particles centrifuge > 2 min at 11,000 x <i>g</i> and take the supernatant for further use.
Low ratio A_{260}/A_{230}	<i>Carry-over of chaotropic salts</i> Refer to detailed troubleshooting "Suboptimal performance of DNA in sequencing, restriction, or ligation reactions - Carry-over of chaotropic salts" and see section 2.5 for detailed information.

7.2 Ordering information

Product	REF	Pack of
NucleoSpin® Gel and PCR Clean-up	740609.10/.50/.250	10/50/250
NucleoSpin® Gel and PCR Clean-up Midi	740986.20	20
NucleoSpin® Gel and PCR Clean-up Maxi	740610.20	20
Buffer NT1	740305.120	200 mL
Buffer NTB	740595.150	150 mL
Buffer NTC	740654.100	125 mL
Buffer NT3 Concentrate (for 125 mL Buffer NT3)	740598	25 mL
Collection Tubes (2 mL)	740600	1000
NucleoTrap®	740584.10/.50/.250	10/50/250
NucleoTrap®CR	740587.10/.50/.250	10/50/250

Visit ***www.mn-net.com*** for more detailed product information.

7.3 References

Vogelstein B., and D. Gillespie. 1979. Preparative and analytical purification of DNA from agarose. Proc. Natl. Acad. Sci. USA 76: 615-619.

7.4 Product use restriction / warranty

NucleoSpin® Gel and PCR Clean-up Midi and NucleoSpin® Gel and PCR Clean-up Maxi kit components are intended, developed, designed, and sold FOR RESEARCH PURPOSES ONLY, except, however, any other function of the product being expressly described in original MACHEREY-NAGEL product leaflets.

MACHEREY-NAGEL products are intended for GENERAL LABORATORY USE ONLY! MACHEREY-NAGEL products are suited for QUALIFIED PERSONNEL ONLY! MACHEREY-NAGEL products shall in any event only be used wearing adequate PROTECTIVE CLOTHING. For detailed information please refer to the respective Material Safety Data Sheet of the product! MACHEREY-NAGEL products shall exclusively be used in an ADEQUATE TEST ENVIRONMENT. MACHEREY-NAGEL does not assume any responsibility for damages due to improper application of our products in other fields of application. Application on the human body is STRICTLY FORBIDDEN. The respective user is liable for any and all damages resulting from such application.

DNA/RNA/PROTEIN purification products of MACHEREY-NAGEL are suitable for IN-VITRO-USES ONLY!

ONLY MACHEREY-NAGEL products specially labeled as IVD are also suitable for IN-VITRO-diagnostic use. Please pay attention to the package of the product. IN-VITRO-diagnostic products are expressly marked as IVD on the packaging.

IF THERE IS NO IVD SIGN, THE PRODUCT SHALL NOT BE SUITABLE FOR IN-VITRO-DIAGNOSTIC USE!

ALL OTHER PRODUCTS NOT LABELED AS IVD ARE NOT SUITED FOR ANY CLINICAL USE (INCLUDING, BUT NOT LIMITED TO DIAGNOSTIC, THERAPEUTIC AND/OR PROGNOSTIC USE).

No claim or representations is intended for its use to identify any specific organism or for clinical use (included, but not limited to diagnostic, prognostic, therapeutic, or blood banking). It is rather in the responsibility of the user or - in any case of resale of the products - in the responsibility of the reseller to inspect and assure the use of the DNA/RNA/protein purification products of MACHEREY-NAGEL for a well-defined and specific application.

MACHEREY-NAGEL shall only be responsible for the product specifications and the performance range of MN products according to the specifications of in-house quality control, product documentation and marketing material.

This MACHEREY-NAGEL product is shipped with documentation stating specifications and other technical information. MACHEREY-NAGEL warrants to meet the stated specifications. MACHEREY-NAGEL's sole obligation and the customer's sole remedy is limited to replacement of products free of charge in the event products fail to perform as warranted. Supplementary reference is made to the general business terms and conditions of MACHEREY-NAGEL, which are printed on the price list. Please contact us if you wish to get an extra copy.

There is no warranty for and MACHEREY-NAGEL is not liable for damages or defects arising in shipping and handling (transport insurance for customers excluded), or

out of accident or improper or abnormal use of this product; defects in products or components not manufactured by MACHEREY-NAGEL, or damages resulting from such non-MACHEREY-NAGEL components or products.

MACHEREY-NAGEL makes no other warranty of any kind whatsoever, and SPECIFICALLY DISCLAIMS AND EXCLUDES ALL OTHER WARRANTIES OF ANY KIND OR NATURE WHATSOEVER, DIRECTLY OR INDIRECTLY, EXPRESS OR IMPLIED, INCLUDING, WITHOUT LIMITATION, AS TO THE SUITABILITY, REPRODUCTIVITY, DURABILITY, FITNESS FOR A PARTICULAR PURPOSE OR USE, MERCHANTABILITY, CONDITION, OR ANY OTHER MATTER WITH RESPECT TO MACHEREY-NAGEL PRODUCTS.

In no event shall MACHEREY-NAGEL be liable for claims for any other damages, whether direct, indirect, incidental, compensatory, foreseeable, consequential, or special (including but not limited to loss of use, revenue or profit), whether based upon warranty, contract, tort (including negligence) or strict liability arising in connection with the sale or the failure of MACHEREY-NAGEL products to perform in accordance with the stated specifications. This warranty is exclusive and MACHEREY-NAGEL makes no other warranty expressed or implied.

The warranty provided herein and the data, specifications and descriptions of this MACHEREY-NAGEL product appearing in MACHEREY-NAGEL published catalogues and product literature are MACHEREY-NAGEL's sole representations concerning the product and warranty. No other statements or representations, written or oral, by MACHEREY-NAGEL's employees, agent or representatives, except written statements signed by a duly authorized officer of MACHEREY-NAGEL are authorized; they should not be relied upon by the customer and are not a part of the contract of sale or of this warranty.

Product claims are subject to change. Therefore please contact our Technical Service Team for the most up-to-date information on MACHEREY-NAGEL products. You may also contact your local distributor for general scientific information. Applications mentioned in MACHEREY-NAGEL literature are provided for informational purposes only. MACHEREY-NAGEL does not warrant that all applications have been tested in MACHEREY-NAGEL laboratories using MACHEREY-NAGEL products. MACHEREY-NAGEL does not warrant the correctness of any of those applications.

Last updated: 07 / 2010, Rev. 03

Please contact:

MACHEREY-NAGEL GmbH & Co. KG

Tel.: +49 (0) 24 21 969 270

e-mail: tech-bio@mn-net.com

Trademarks:

NanoDrop® is a registered trademark of NanoDrop Technologies, Inc.

NucleoSpin® is a registered trademark of MACHEREY-NAGEL GmbH & Co. KG

NucleoTrap® is a registered trademark of MACHEREY-NAGEL GmbH & Co. KG

All used names and denotations can be brands, trademarks, or registered labels of their respective owner – also if they are not special denotation. To mention products and brands is only a kind of information (i.e., it does not offend against trademarks and brands and can not be seen as a kind of recommendation or assessment). Regarding these products or services we can not grant any guarantees regarding selection, efficiency, or operation.



Plasmid DNA

Clean up

RNA

DNA

Viral RNA and DNA

Protein

High throughput

Accessories

Auxiliary tools



www.mn-net.com

MACHEREY-NAGEL



MACHEREY-NAGEL GmbH & Co. KG · Neumann-Neander-Str. 6-8 · 52355 Düren · Germany

DE/International:

Tel.: +49 24 21 969-0
Fax: +49 24 21 969-199
E-mail: info@mn-net.com

CH:

Tel.: +41 62 388 55 00
Fax: +41 62 388 55 05
E-mail: sales-ch@mn-net.com

FR:

Tel.: +33 388 68 22 68
Fax: +33 388 51 76 88
E-mail: sales-fr@mn-net.com

US:

Tel.: +1 484 821 0984
Fax: +1 484 821 1272
E-mail: sales-us@mn-net.com



A040801/1290.1